



waitkus360°

BIOCOMPATIBILITY CERTIFICATE



HP Inc.

HP 3D600/3D700 Fusing and Detailing Agents and HP 3D HR PA12 USP Class I-VI and FDA Intact Skin Surface Devices Certification

HP 3D600/3D700 Fusing and Detailing Agents and HP 3D HR PA 12 have met the requirements of USP Class I-VI and US FDA's guidance for Intact Skin Surface Devices. This conclusion is based on following tests and guidelines used:

1. USP Class I-VI including irritation, acute systemic toxicity, and implantation
2. Cytotoxicity – ISO 10993-5, Biological evaluation of medical devices – part 5: Tests for in vitro cytotoxicity, and
3. Sensitization – ISO 10993-10, Biological evaluation of medical devices – Part 10: Tests for irritation and skin sensitization.

HP believes that the testing referred to above is representative of parts produced with HP 3D600/3D700 Fusing and Detailing Agents and fresh HP 3D HR PA 12 powder¹ on the HP Jet Fusion 3D 3200 and 4200 printers. Based on these results, HP expects that similar articles made from these materials, under similar conditions will meet the compliance requirements of USP Class I-VI and FDA's guidance for Intact Skin Surface Devices.

It is the responsibility of each customer to determine that its use of HP 3D600/3D700 Fusing and Detailing Agents and HP 3D HR PA 12 powder is safe and technically suitable to the customer's intended applications and consistent with the relevant regulatory requirements (including FDA requirements) applicable to the customer's final product. Customers should conduct their own testing to ensure that this is the case. Results may vary if the testing is performed under different conditions than those existing at HP's laboratories at testing time and those that applied for the purposes of the biocompatibility tests as referenced above. Because of possible changes in the relevant industry standards, FDA guidance, and other legal or regulatory requirements, as well as possible changes in HP 3D600/3D700 Fusing and Detailing Agents and HP 3D HR PA 12 powder, HP cannot guarantee that the status of HP 3D600/3D700 Fusing and Detailing Agents and HP 3D HR PA12 powder will remain unchanged or that it will qualify for USP Class I-VI Certification and or comply with FDA's guidance for Intact Skin Surface Devices in any particular use.

For additional information about HP 3D600 Fusing and Detailing Agents and HP 3D HR PA 12, please contact our HP 3D Printing Materials team at 3dmaterials@hp.com.

¹ Testing performed with 100% fresh powder



waitkus360°

© Copyright 2017 HP Development Company, L.P.

The only warranties for HP products and services are set forth in the express warranty statements accompanying such products and services. Nothing herein should be construed as constituting an additional warranty. HP shall not be liable for technical or editorial errors or omissions contained herein.

4AA7-0331EEW, June 2017

This is an HP Indigo digital print.

